

1. Identification of the Substance and Company

Revital-Ox™ RESERT™ R60 Solution Test Strip

Product No. PCC053, PCC054

SDS No. S296US

Prepared by: M. Ebers

asksteris_msds@steris.com

Date Created: 07/16/2015

Date Revised: 02/19/2018

NFPA 704 HAZARD RATING:

HEALTH: 0

FIRE: 0

REACTIVITY: 0

Date Reviewed: NA

STERIS Corporation, P.O. Box 147, St. Louis, Mo 63166, US
 Emergency Telephone No. 1-314-535-1395 (STERIS); 1-800-424-9300 (CHEMTREC)
 Telephone Number for information: 1-800-548-4873 (Customer Service – Healthcare Products)

STERIS Ireland Limited, IDA Business and Technology Park, Tullamore
 County Offaly, R35 X865, Ireland.

Emergency Phone No: +44 (0) 1895 622 639

Product/Technical Information Phone No: +44 (0) 116 276 8636

PRODUCT NAME:	PRODUCT NUMBER:
Revital-Ox™ RESERT™ R60 Solution Test Strip (60 strips)	PCC053
Revital-Ox™ RESERT™ R60 Solution Test Strip (2 X 60 strips)	PCC054
SHIPPING INFORMATION:	NOT RESTRICTED

These products are NOT hazardous products according to the OSHA Hazard Communication Standard, 29 CFR 1910.1200.

The Revital-Ox™ RESERT™ R60 Solution Test Strip is free of latex and natural dry rubber products.

The Revital-Ox™ RESERT™ R60 Solution Test Strip does not contain any components derived from animal origin, does not come into contact with materials of animal origin and therefore complies with the "Note for guidance of minimizing the risk of transmitting animal spongiform encephalopathy agents via human Veterinary medicinal products (EMEA/410/01 Rev. 2-October 2003)."

The Revital-Ox™ RESERT™ R60 Solution Test Strip is manufactured in the absence of lead and other heavy metals.

In the case of accidental exposure, take the following precautionary measures:

EYES: Immediately flush eyes with plenty of water for 15 minutes. Get medical attention if irritation persists.

SKIN: Immediately wash with soap and water. Get medical attention if irritation develops.

INGESTION: Give water to drink and get medical advice.

INHALATION: Remove to fresh air. Get medical attention for any breathing difficulty.

Chemical Monitoring Strips Handling and Disposal:

Handling and Storage Precautions: Keep cap tightly closed. Store at 6°C (43°F) to 30°C (86°F) and 30-80% humidity. Do not freeze. Protect strips from exposure to light, heat and moisture. Protect strips from exposure to oxidizing agents, strong acids/alkalis and detergents.

Waste Disposal Methods: Dispose of used/expired chemical indicators and boxes in trash receptacle in accordance with federal, state, and local regulations.